

Phone: (352) 293-2810

Web: www.dynamicdiagnostic.com

Dear Patient,

This oximetry test has been ordered by your physician to monitor and record your blood oxygen saturation level and heart rate. This test is used to determine the need for home oxygen therapy or if further sleep testing is needed. Please wear the unit for at least four (4) hours tonight and call Advanced Diagnostic if you have any questions or testing issues this evening.

Thank you,

Advanced Diagnostic Solutions
Management & Support Team

_____ Oximetry Test Courier _____

DME: _____

Phone: _____

920 M Plus Pulse Oximeter Description



Follow these simple to understand instructions to complete this Oximetry Test.

1. Start – Read Below

Physician's Orders are to test on:

- Room Air (**NO OXYGEN**)
- Oxygen at ____ Liters per Minute
- CPAP BIPAP APAP
- Oxygen at ____ Liters per Minute with your CPAP or BIPAP

2. Probe Connection

Secure the probe at the top of the device.



3. Probe Placement



Wire should be on top of your finger. A piece of tape or Band-Aid can be used to ensure probe stays attached tonight.

4. Turn Unit On

Press & Hold the power button for 2-3 seconds until the display lights up.



Power Button
(On & Off Button)

5. Time for Sleep



If for any reason you encounter issues tonight, please read below to troubleshoot an issue. For further assistance please call Oximetry Support.

Patient Support Lines: (352) 293-2810

6. End Test – Turn Unit Off

Press & hold the power button for 2-3 seconds until the display shuts off.



Power Button
(On & Off Button)

Troubleshooting: Problems or Issues with the Pulse Oximeter?

Battery Indicator Blinking? Complete the test and notify to courier picking up the equipment. Do not attempt to change the batteries as this can delete the information recorded prior to the battery indicator blinking.

Pulse Indicator Red? Try to reposition the finger probe as it may have disconnected from your finger or the machine. If still an issue, notify our office in the morning or call the local courier and notify them.

Blank Display? If there are no red numbers on the display, simply hit the power button to try and turn the unit back on. If this fails please contact our office in the morning to let us know the test failed.